



ENVIRONMENT
AGENCY

PRODUCT CONFORMITY CERTIFICATE

This is to certify that the

Protoc 300 TOC Analyser

manufactured by:

Pollution & Process Monitoring Ltd

Bourne Enterprise Centre

Borough Green

Sevenoaks

TN15 8DG

Kent

UK

has been assessed by Sira Certification Service
and for the conditions stated on this certificate complies with:

**MCERTS Performance Standards for Continuous Water
Monitoring Systems, Version 2.1 (July 2006)**

Certification Ranges :

TOC 0 to 20 mg/L
 0 to 4000 mg/L

Project No: 674/0112
Certificate No: Sira MC 060077/00
Initial Certification: 11 August 2006
This Certificate Issued 11 August 2006
Renewal Date: 10 August 2011

Managing Director

MCERTS is operated on behalf of the Environment Agency by

Sira Certification Service

12 Acorn Industrial Park, Crayford Road, Crayford
Dartford, Kent, UK, DA1 4AL

Tel: 01322 520500 Fax: 01322 520501

This certificate may only be reproduced in its entirety and without change



ENVIRONMENT
AGENCY

Approved Site Application

On the basis of the assessment this instrument is considered suitable for use on treated wastewater and untreated wastewater applications.

Any potential user should ensure, in consultation with the manufacturer, that the product is suitable for the process on which it will be installed.

Basis of Certification

This certification is based on the following Test Report(s) and on Sira's assessment and ongoing surveillance of the product and the manufacturing process:

Environment Agency Warrington	Report Ref: EA-MCERTS-001 dated December 2005
Environment Agency Warrington	Report Ref: EA-MCERTS-002 dated December 2005
Environment Agency Warrington	Report Ref: EA-MCERTS-003 dated April 2006
Sira Laboratory	Report Ref: N51D12061B dated 26/07/05

Product Certified

This certificate applies to all instruments fitted with software version 05 onwards (i.e. serial number 29041324 onwards)

Certificate No: Sira MC 060077/00
This Certificate Issued: 11 August 2006



ENVIRONMENT
AGENCY

Certified Performance

The instrument was evaluated for use under the following conditions:

Ambient Temperature Range: +2°C to +40°C

Results are expressed as expanded uncertainties as a percentage of the certification range 0 to 20mg/L, unless otherwise stated.

Test	Results expressed as expanded uncertainties as a percentage of the certification range				Other results	MCERTS specification
	<0.5	<1	<2	<4		
Combined performance characteristic, U_C					2.46%	10%
Mean error x					-3.70%	10%
Mean error x (0 to 4000mg/L range)					7.1%	10%
Linearity, X_L		-0.92			Note 1	5%
Linearity, X_L (0 to 4000mg/L range)				-3.41	Note 1	5%
Repeatability, U_r		0.64				5%
Repeatability, U_r (0 to 4000mg/L range)			1.24			5%
Drift, X_D				-2.91		5%
Response time in the lab (0 to 20mg/L range)					4min 49s	To be reported
Interference due to inorganic carbon, X_{IN}	0.89					3% span
Analogue output signal impedance, X_{OP}		-0.69				2.5%
Power supply, X_V	-0.26					2.5%
Loss of supply					Pass	Data shall be correctly retained
Sample flow rate, X_{SQ}					See note 2	2.5%
Sample pressure, X_{SP}					See note 2	2.5%
Sample temperature, X_{ST}	-0.2					5%
Ambient temperature, X_T					4.10%	5%
Relative humidity, X_{RH}				3.03		5%

Certificate No: Sira MC 060077/00
This Certificate Issued: 11 August 2006

This certificate may only be reproduced in its entirety and without change



**ENVIRONMENT
AGENCY**

Test	Results expressed as expanded uncertainties as a percentage of the certification range				Other results	MCERTS* specification
	<0.5	<1	<2	<4		
Up-time					97.6%	To be reported
Response time in the field: Initial at average reading of 4.3mg/L (start of field test)					5 min	To be reported
Final at average reading of 4.06mg/L (end of field test)					8min 30s	To be reported
Response time in the field: Initial at average reading of 15mg/L (start of field test)					6min 15s	To be reported
Final at average reading of 16.88mg/L (end of field test)					6min 45s	To be reported
Field test error					90.5%	To be reported

Note 1: Method of calculation is terminal based linearity.

Note 2: Tests not applicable as a pump controls the sample flow-rate and sample pressure.

Note 3: Field test: The Protoc 300 analyser was assessed on the basis of a three month field trial installed on an urban river application.

Certificate No: Sira MC 060077/00
This Certificate Issued: 11 August 2006

This certificate may only be reproduced in its entirety and without change



ENVIRONMENT
AGENCY

Description:

The Protoc 300 TOC analyser measures a sample stream continuously for organic carbon content. Sample is mixed with an acidified reagent containing sodium persulfate. The mixture of sample and reagent is sparged with a carrier gas (usually compressed air) to remove inorganic carbon and dissolved carbon dioxide. The remaining liquid, which now contains only organic carbon compounds, is pumped into the reaction vessel. A strong source of ultra-violet light promotes an oxidation reaction between the organic and persulfate, producing carbon dioxide. A stream of carrier gas diffusing at the bottom of the reaction vessel, transfers the carbon dioxide through a gas drying system into a non-dispersive infrared detector. The amount of carbon dioxide produced is directly proportional to the organic carbon content of the sample over a specific measuring range determined by the application.

The analyser may be set up to measure total organic carbon (TOC) or total carbon (TC). In the latter case the sparging of the sample is omitted.

The analyser is calibrated by passing a standard test solution through the analyser. The standard contains a known level of carbon and the detector is adjusted to display the correct reading. Solenoids are used to switch either the sample or the standard solution through the analyser.

General Notes

1. This certificate is based upon the product tested. The Manufacturer is responsible for ensuring that on-going production complies with the standard(s) and performance criteria defined in this Certificate. The Manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management system shall be subject to regular surveillance according to 'Regulations Applicable to the Holders of Sira Certificates'. The design of the product certified is defined in the Sira Design Schedule for certificate No. Sira MC 060077/00.
2. If certified product is found not to comply, Sira Certification Service should be notified immediately at the address shown on this certificate.
3. The Certification Marks that can be applied to the product or used in publicity material are defined in 'Regulations Applicable to the Holders of Sira Certificates'.
4. This document remains the property of Sira and shall be returned when requested by the company.

Certificate No: Sira MC 060077/00
This Certificate Issued: 11 August 2006